A Nationally Accredited Agency

Certificate/Report of Evaluation

Manufacturer:	Fluke	Test Results:	Passed
Model:	77	Date Checked:	8/25/04
Description:	Multimeter II	Due Date:	8/25/05
Serial #:	66710035	Received:	In Tolerance
Technician:	Bruce Smith	Returned:	In Tolerance
Report #:	4-66710035		

Reference Used:

Manufacturer	Model	Description	Cal Date	Due Date
Fluke (SN7733010)	45	Multimeter	10-10-2003	10-10-2004

Test Data:

Test Points (BAC DataMaster #921012)	Reference	Test Meter	Acceptable Range
TP1 and TP5 (Sample Chamber Board)	2.40 V_{DC}	2.395	+/- 0.3% (2.39 – 2.41 V_{DC})
TP5 and TP4 (Sample Chamber Board)	3.031 V_{DC}	3.026	+/- 0.3% (3.02 – 3.04 V_{DC})
TP4 and TP1 (Detector Board) (or bottom of R45)	120.00 V_{DC}	119.7	+/- 0.3% (119.6 – 120.4 V_{DC})

COMMENT: AS FOUND DATA IN TOLERANCE. NO ADJUSTMENT REQUIRED.