A Nationally Accredited Agency

Certificate/Report of Evaluation

Manufacturer:	Fluke	Test Results:	Passed
Model:	77	Date Checked:	8/10/06
Description:	Multimeter III	Due Date:	8/28/07
Serial #:	75350847	Received:	InTolerance
Technician:	Bruce Smith	Returned:	In Tolerance
Report #:	6-75350847		

Reference Used:

Manufacturer	Model	Description	Cal Date	DueDate
Fluke (SN7733010)	45	Multimeter	9-22-05	9-22-06

Test Data:

Test Points (BAC DataMaster #950387)	Reference	Test Meter	Acceptable Range
TP1 and TP5 (Sample Chamber Board)	2.40 V_{DC}	2.398	+/- 0.3% (2.39 – 2.41 V_{DC})
TP5 and TP4 (Sample Chamber Board)	3.031VDC	3.029	+/- 0.3% (3.02-3.04V_{DC})
TP4 and TP1 (Detector Board) (or bottom of R45)	120.00 V_{DC}	119.8	+/- 0.3%(119.6-120.4V_{DC})

COMMENT : AS FOUND DATA IN TOLERANCE. NO ADJUSTMENT REQUIRED.